



Negative Control Mouse IgG1

Negative Control

Control Number: 901-NC490-030308

ISO
9001:2000
CERTIFIED

Catalog Number: NC490 AA, H
Description: 6.0, 25 ml, prediluted

Intended Use:
For In Vitro Diagnostic Use

Summary and Explanation:
The MOPC-21 tumor line that produces mouse IgG1/kappa is a mineral oil-based plasmacytoma carried in BALB/c mice. The IgG1 preparation has been purified from mouse ascites fluid. An IgG1 negative control is usually matched at the same concentration as the primary antibody (IgG1 isotype). It can also be used on an automated staining system.

Principle of Procedure:
Antigen detection, in tissues and cells, is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a universal, affinity-purified, secondary antibody is added to bind to the primary antibody. An enzyme label is then added to bind to the secondary antibody; this detection of the bound antibody is evidenced by a colorimetric reaction.

Source: Purified Mouse Immunoglobulin's

Isotype: IgG₁/kappa

Total Protein Concentration: ~1 µg/ml

Known Applications:
Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Purified mouse IgG₁/kappa in buffer with protein carrier preservative

Storage and Stability:
Store at 2°C to 8°C. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations

Peroxide Block:
If using an HRP system, block for 5 minutes with BIOCARE's PEROXIDAZED 1.

Protein Block:
Incubate for 10-15 minutes at RT with BIOCARE's Background Sniper.

Pretreatment Protocol:
Please refer to the respective primary antibody datasheet for recommended pretreatment solution and protocol.

Negative Control:
Use a nonspecific negative reagent control in place of the primary antibody with a section of each patient specimen to evaluate nonspecific staining and allow better interpretation of specific staining at the antigen site.

Link: Incubate for 10 minutes at RT with a link

Label: Incubate for 10 minutes at RT with a label.

Chromogen:
Incubate for 5 minutes at RT when using BIOCARE's DAB. - OR - Incubate for 10 minutes at RT when using BIOCARE's Vulcan Fast Red.

Technical Note:
This antibody has been standardized with BIOCARE's 4 plus detection system. It can also be used on an automated staining system and with other BIOCARE polymer detection kits.

Performance Characteristics:
The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of BIOCARE products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Quality Control:
Refer NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information on Tissue Controls.

Precautions:
This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976)
Specimens, before and after fixation and all materials exposed to them, should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. Microbial contamination of reagents may result in an increase in nonspecific staining. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change. The MSDS is available upon request.

Troubleshooting:
Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact BIOCARE's Technical Support at 1-800-542-2002.

Limitations and Warranty:
There are no warranties, expressed or implied, which extend beyond this description. BIOCARE is not liable for property damage, personal injury, or economic loss caused by this product.

